
MARKET ANALYSIS

ENG.ARAS KAMAL MESLEM

PENTACAM

- Three Models
- Normal
- HR
- AXL





PENTACAM WINNING POINTS

- Market Leader
- Good Installment Base
- Good Clinical studies
- Good Marketing
- Considered the Golden standard of Anterior Chamber Imaging Based on Scheimpflug camera technology
- Connected to Alcon / WaveLight Exciemr Laser


PENTACAM CONS

- Only Scheimpflug camera no Placido Disc





PENTACAM CONS



Patient's data

Ce: [redacted] Fr: [redacted]
Birthdate: 03/08/1934
Identification code: 0000396306
Exam date and time: 10/11/2009 10.45
Group: Post-Cataract
Acquisition date: 10/11/2009 10.50.17

Summary Indices

Horizontal Visible Iris Diameter
HVID = 12.00 mm
Pupil (Topographic) +
r = 0.18 mm @ 191°
Ø = 2.89 mm
Thinnest location +
r = 0.60 mm @ 244°
Thk = 593 μm
Apex x
r = 0.40 mm @ 165°
Thk = 596 μm Curv = 45.89 D
Anterior chamber
Volume = 156 mm³
Iridocorneal angle = 56°

K readings (Front)

Sim-K
n0 = 1, n1 = 1.3375
Sim-K
K1 = 43.51 D @ 170°
K2 = 43.71 D @ 80°
Avg = 43.61 D
Cyl = -0.21 D Ax170°

PENTACAM CONS

- Three Models no upgrades
- AXL very expensive
- Poor results of Biometry compared to Standard market instruments
- The machine is relatively cheap
- Pay for the software

MODULAR SOFTWARE

- General Overview
- Fast Screening Report
- Topography and Elevation Maps of the Anterior and Posterior Corneal Surface
- Pachymetry Maps, absolute and relative
- Scheimpflug Image Overview
- 3D Anterior chamber analysis
- Anterior segment tomography
- Topographical Keratoconus Classification (TKC)
- Belin ABCD Keratoconus Staging

MODULAR SOFTWARE

- Belin ABCD Progression Display
- Iris Image and HWTW
- 4 Maps Refractive
- Compare 2 Exams
- Belin/Ambrósio Enhanced Ectasia Display
- Contact Lens Fitting

REFRACTIVE PACKAGE

- Freely selectable reference bodies for elevation maps
- Corneal Optical Densitometry
- Corneal Rings
- Corneal analysis for refractive surgeons (corneal optical densitometry)
- Corneal thickness progression analysis for early keratoconus detection
- 4 Maps Selectable
- Show 2 Exams
- Compare 4 Exams
- Side-by-side comparison of topometric and pachymetric data
- Fourier Analysis

CATARACT PACKAGE

- Cataract Pre-OP Display for premium IOL selection and assessment of the optical properties of the entire cornea
- True Net Power map (TNP)
- Total Corneal Refractive Power map (TCRP)
- Corneal Power Distribution
- Zernike Analysis including normative corneal wavefront data
- PNS and 3D Cataract Analysis
- Show 2 Exams
- Measurements in the Scheimpflug images

CATARACT PACKAGE

- Automatic calculation of the anterior chamber angle in 360°, measurement based on Scheimpflug images
- 4 Maps Chamber
- Compare 4 Exams
- Comparative analysis of topometric and pachymetric data
- 4 Maps Topometric
- Anterior Chamber Depth Map
- 3D anterior chamber analysis
- Normative Corneal Wavefront data including anterior, posterior and total corneal data

CATARACT PACKAGE

- Compare 4 Exams
- Comparative analysis of topometric and pachymetric data
- Automatic calculation of the anterior chamber angle in 360° , measurement based on Scheimpflug images
- 4 Maps Chamber
- 4 Maps Topometric
- Anterior Chamber Depth Map
- Measurements in the Scheimpflug images
- 3D anterior chamber analysis
- Normative Corneal Wavefront data including anterior, posterior and total corneal data

COST

- Total Cost is 30% is higher than the Sirius










www.alamy.com - C06NA7


KERATOCONUS SCREENING

- Belin/Ambrósio Enhanced Ectasia Display
- Based on Anterior & Posterior Elevation and the Pachymetry

SIRIUS KERATOCONUS SCREENING

- Build on the Artificial Intelligence Concept
- The first study conducted with several thousands of Patients
- The System developes it self with thousands of cases
- Uses multiple factors and several maps to classify Keratoconus ;no only the posterior elevation and Pachymetry





IOL CALCULATION POST LASIK

Holladay Report and Holladay EKR65 Detail Report

Poor not accurate

Pentacam Keratometric Values Unreliable for IOL Power Calculation After Refractive Surgery

Pentacam Keratometric Values Unreliable for IOL Power Calculation After Refractive Surgery

To the Editor:

We read with interest the article by Gonen et al,¹ which appeared in the August 2012 issue of the *Journal of Refractive Surgery*, on their study comparing keratometric data from 200 patients evaluated for refractive surgery using four different devices—an automated keratometer (KR 8800 [Topcon, Tokyo, Japan]), two Placido-based computerized topography systems (Dicon CT 200 [Vismed Inc, San Diego, California] and Allegro Topolyzer [WaveLight Inc, Sterling, Virginia]), and a Scheimpflug analyzer (Pentacam [Oculus Optikgeräte GmbH, Wetzlar, Germany]). The study revealed that although the mean simulated keratometry (Sim K) obtained with the Pentacam did not differ significantly from the other three devices, the Pentacam gave rise to the greatest number of extreme outliers (>2.00 -diopter [D] difference) on Bland-Altman plots. In addition, the 95% limits of agreement for the Pentacam Sim K value compared to the other devices was approximately 1.35 D.

the mean absolute error (MAE) was 0.92 ± 0.74 D, with 33% of eyes exhibiting an MAE of >1.00 D.

Other studies reflect these findings.²⁻⁴ Interestingly, Kim et al⁵ used Pentacam true net corneal power to calculate the IOL power for eyes requiring cataract surgery following keratorefractive surgery. Although postoperative refraction was ± 0.50 D of intended in 70% of eyes (21/30), conversely, $>25\%$ of eyes had ≥ 0.75 D MAE compared to the desired refractive outcome.

In a climate of refractive excellence, where surgeons target ± 0.25 D of refractive error, the postoperative PRK and LASIK keratometric values provided by the Pentacam no longer appear acceptable in the calculation of IOL power. With increasing numbers of “baby-boomers” who have already undergone refractive surgery looking for equal excellence in cataract outcomes, surgeons must continue the search for accurate and reproducible keratometric values following corneal refractive surgery.

Jennifer C. Fan Gaskin, MBChB
Chi-Ying Chou, MBChB
Charles N.J. McGhee, PhD, FRCOphth
Auckland, New Zealand

gerate GmbH, Wetzlar, Germany]]. The study revealed that although the mean simulated keratometry (Sim K) obtained with the Pentacam did not differ significantly from the other three devices, the Pentacam gave rise to the greatest number of extreme outliers (>2.00 -diopter [D] difference) on Bland-Altman plots. In addition, the


SIRIUS IOL CALCULATION

- Corneal ray tracing versus simulated keratometry for estimating corneal power changes after excimer laser surgery

METHODS: This multicenter retrospective study comprised patients who had myopic or hyperopic excimer laser refractive surgery. Preoperatively and postoperatively, 2 corneal power measurements—simulated keratometry (K) and mean pupil power—were obtained. The mean pupil power was the corneal power calculated over the entrance pupil by ray tracing through the anterior and posterior corneal surfaces using Snell's law. Agreement between the refractive and corneal power change was analyzed according to Bland and Altman. Regression analysis and Bland-Altman plots were used to evaluate agreement between measurements.

RESULTS: The study evaluated 72 eyes (54 patients). The difference between the postoperative and preoperative simulated K values underestimated the refractive change after myopic correction and overestimated it after hyperopic correction. Agreement between simulated K changes and refractive changes was poor, especially for higher amounts of correction. A proportional bias was detected ($r = -0.77$; $P < .0001$), and the 95% limits of agreement (LoA) were $-0.15 - 0.14 \times \pm 0.62$ diopters (D). The difference between the postoperative and preoperative mean pupil power showed an excellent correlation with the refractive change ($r^2 = 0.98$). The mean pupil power did not overestimate or underestimate the refractive change. The 95% LoA ranged between -0.97 D and $+0.56$ D.

CONCLUSION: Corneal ray tracing accurately measured corneal power changes after excimer laser refractive surgery.



WHAT SIRIUS DOES NOT HAVE

- Biometry
- Fourier Analysis
- PNS & 3D cataract Analysis
- 3D pIOL Simulation and Aging Prediction

Model Line-Up

"OCULUS" Optikgeraete GmbH [DE] | https://www.pentacam.com/int/ophthalmologist-surgeon-without-pentacam/models/model-line-up.html

I AM ...
START
TECHNOLOGY
MODELS
PENTACAM® LIVE
PUBLICATIONS
ABOUT US
CONTACT

Scheimpflug Camera	Pentacam®	Pentacam® HR	Pentacam® AXL
Camera	digital CCD camera	digital CCD camera	digital CCD camera
Light source	blue LEDs (475 nm UV-free)	blue LEDs (475 nm UV-free)	blue LEDs (475 nm UV-free)
Processor	DSP with 400m operations/s	DSP with 400m operations/s	DSP with 400m operations/s
Speed	50 images in 2 seconds ¹⁾	100 images in 2 seconds ²⁾	100 images in 2 seconds ²⁾
Measurement range	Pentacam®	Pentacam® HR	Pentacam® AXL
Curvature	3 to 38 mm 9 to 99 D	3 to 38 mm 9 to 99 D	3 to 38 mm 9 to 99 D
Precision	± 0.2 D	± 0.1 D	± 0.1 D
Reproducibility	± 0.2 D	± 0.1 D	± 0.1 D
Operating distance	80 mm (3.1 in)	80 mm (3.1 in)	80 mm
Technical specifications	Pentacam®	Pentacam® HR	Pentacam® AXL
Dimensions (W x D x H)	275 x 320-400 x 500-530 mm (10.8 x 12.6 - 15.7 x 19.7 - 20.9 in)	275 x 320-400 x 500-530 mm (10.8 x 12.6 - 15.7 x 19.7 - 20.9 in)	275 x 320-400 x 500-530 mm (10.8 x 12.6 - 15.7 x 19.7 - 20.9 in)
Weight	10.1 kg (22.3 lbs)	10.6 kg (23.4 lbs)	11.2 kg (24.7 lbs)
Power input, max.	35 W	42 W	37.4 W
Recommended computer specifications	CPU Intel Core i5-6600, HDD 1 TB, RAM 8 GB, MS Windows® 10 Pro, VESA, USB interface	CPU Intel Core i5-6600, HDD 1 TB, RAM 8 GB, MS Windows® 10 Pro, VESA, USB interface	CPU Intel Core i5-6600, HDD 1 TB, RAM 8 GB, MS Windows® 10 Pro, VESA, USB interface

Model Line-Up.html
csm_pentacam_84....png
pentacam-hr.png
pentacam-axl.png
Show all

ENG US
9:50 AM 4/6/2018

PENTACAM WHAT DOES NOT HAVE

- Simulation quality of vision
- pupillometry
- Tear film analysis
- meibografy
- Integration with other instruments like Wavefront analyzer and SL
- Advanced Alemetric
- Front Gaussian
- Advanced Pachymetry

GALILEI

**FOR REFRACTIVE AND
CATARACT SURGERY**

GALILEI G4

REACHING A NEW LEVEL IN CORNEAL
TOPOGRAPHY AND TOMOGRAPHY



GALILEI

THE NEW GALILEI G6

REACHING A NEW LEVEL IN
BIOMETRY WITH HD TOPOGRAPHY
AND TOMOGRAPHY


**„THE MOST COMPLETE ALL-IN-ONE SOLUTION,
FROM REFRACTIVE TO CATARACT SURGERY“**


The GALILEI G6 Lens Professional is pending FDA approval and is not available for sale in the US. For some other countries, availability may be restricted due to local regulatory requirements. Please contact Zeimer for details.





GALILEI

- Placido Disk
- Dual Scheimpflug camera





DUAL CAMERA



GALILEI CONS

- Not easy to install , needs certified engineer to installed
- Normally you will end with broken fiber optic
- Adding dual camera means to many mechanical parts , slow, noisy , and the whole system moves when you start measuring
- Screening monitor , more connections and more problems , not easy to replce needs to be send to the factory

GALILEI CONS


- Not very sold in the market no large instalments
- Software not modular , but the it is poor only the basic functions
- IOL Calculation Poor

FEATURES DOES NOT HAVE

- Simulation quality of vision
- pupillometry
- Tear film analysis
- meibografy
- Integration with other instruments like Wavefront analyzer and SL
- Advanced Alemetric
- Front Gaussian
- Advanced Pachymetry

ORBISCAN

- Not Scheimpflug camera
- Scanning slit
- each single Slit Image from limbus to limbus, but overlay of the slit images only in the central 5mm zone, 2D-information, add. Placido system
- 40 slit images, each one with 300-350 data points, sum: 12.000 – 14.000 data points



DISADVANTAGES OF THE OBRISCAN

- Long examination time
- Old technology , over 20 years old
- Not reliable Pachymetry
- You have to optimize the Pachymetry with the factor
- No IOL calculation
- No Online support , hard to make service

TOMEY TMS-5

- Small foot print
- Higher speed
- 64scan in 1 sec

TMS-5 CONS

- Using small Placido disc ,theoretically should be cover more of the cornea , but you have to be so close to the eye of the Patient and it s very uncomfortable
- Take picture for the Placido then the camera should turn and position the Scheimpflug ,
Too much mechanics and lose the centration
- Built in PC , if you have any problem of the PC the system will not work
- Connectivity
- Non Online Support

Thank You

